

### **Amendment No.4**

<b><u>To:</u></b>	Interested Bidders
<b><u>From:</u></b>	UNDP Syria
<b><u>Subject:</u></b>	<b>Supply of wheelchairs in Damascus</b>
<b><u>Reference:</u></b>	UNDP-SYR-ITB-020-19
<b><u>Date:</u></b>	18 Apr. 2019

*\* UNDP Syria is answering on the received questions as in the following:*

	<b>Question</b> السؤال	<b>Answer</b> الجواب
<b>1</b>	Item#1, the requirement of adjustable backrest height, approximate range from 30 to 45 cm and tension adjustable backrest, with several tensionable bands. These two requirement won't exist at the same time!!! Can you confirm?.	The two requirements (adjustable backrest height, approximate range from 30 to 45 cm, and tension adjustable backrest with several tensionable bands) are mandatory and both must be present at the same time.
<b>2</b>	Item#1, the requirement of heights 42 and 52 cm to be achieved with the front wheel of 5", this can not be achieved, is it essential?	yes, it is essential. Please note that with the 5 "front wheels you can reach these heights (42-52) using forks with different sizes.
<b>3</b>	As the specification are customized and rarely to find a ready-manufactured wheel chair the same, so the cushions might not be suitable!! Can you accept he ready size manufactured?	Cushion size should match the chair seat size, a difference of about 1 cm for the width and 1 cm for the depth between the measurements of the cushion and those of the wheelchair seat can be accepted.

4	Item#2, the requirement of total weight 13kg, is it a must or another weight is accepted?	The requirement is weigh <u>less than</u> 13 kg. any weigh of under 13 kg and responsive to the other requirements is acceptable. 13 kg and above is not acceptable
5	Item#3 the requirement of rigid frame, is it a must or you can accept alternative?	The rigid frame is mandatory