

## **Requested Clarifications**<sup>4</sup>

Dear prospective bidders,

UNDP Tanzania would like to share additional responses out of requested clarifications from some of you as follows:

**Item 1:** Centrifuge

In your clarifications you have stated that you need a 12tubes x 15ml capacity centrifuge and indicated 8tubes x 50ml tubes under additional rotor. Do you require 2 rotors i.e 12 x 15ml and 8 x 50ml or just one rotor. If one rotor, what capacity?

**\*Response:** Re-sharing again the specifications as per bellow:

Item: Digital Microprocessor based Centrifuge machine with timer | 12 tubes | Bench Top  
Material: Metal Ms Body  
Dimensions: 42 x 34 x 28 cm  
Voltage: 220 v  
RPM: 4500-5000  
Weight: 14.4 Kg  
Capacity: 15 ml x 12 tubes  
Item shape: Square

**Item 11:** Weather station - The below lists different size solar panels, rain gauges, temperature sensors, kindly specify the needed items.

**\*Response:** Weather station- this will be stationed outside the collection centers, appreciate if you could please advise from experience of supplying such equipments what panel sizes are ideal for such environments? (we need basic sensors like the temperature sensors, etc.)

**Item 14:** Grading tables & Item 15. Screening tables

In clarification response 2 you asked us to advise if these are available at furniture stores. They are NOT available. They would have to either be custom made or procured by a specialized company. Basically, suppliers that concentrate on analytical laboratory equipment do not deal in these tables and we will have to procure as a 3rd/4th party is we are even able to. It may not be up to the standard required as well as the price will be very high thus costing the procuring entity unnecessary high charges when they can get these directly from the suppliers that deal with these. In your best interest it really isn't a cost effective or quality effective purchase.

Kindly research and revisit these 2 items in this procurement process for your own benefit.

**\*Response:** Have revisited and concluded that the two items should remain in the list. Thus, therefore proceed to bid for these two items.

UNDP Tanzania wish you all the best.

Best Regards,  
Procurement Unit, UNDP Tanzania