

Date: 10<sup>th</sup> Sep 2020

## ADDENDUM 01

**To:** All Potential Bidders  
**From:** UNDP CYPRUS

**Subject:** UNDP-CYP-RFQ-176-2020 – Supply and Delivery of Refrigerator and other Medical Equipment for (COVID-19) to Nicosia – Cyprus

Reference is made to the UNDP published procurement notice ref: **UNDP-CYP-RFQ-176-2020 – Supply and Delivery of Refrigerator and other Medical Equipment for (COVID-19) to Nicosia – Cyprus** on UNDP website.

Based on bidders clarification and UNDP' Project decision, all interested potential bidders are hereby notified that:

1. The specification for **Mini-Centrifuge/Vortex Devices** are amended to read as follows:

### **Mini-Centrifuge/Vortex devices Model MSC 6000 and FVL 2400 – Or Equivalent**

The hospital lab requires two **MSC 6000** and two **FVL 2400** as per below specifications or equivalent product:

Model	FV-2400	FVL-2400N	MSC-3000	MSC-6000
Centrifugation mode	Fixed, 2800 or 3500* RPM		1000 – 3500 RPM	1000-6000 RPM
Relative centrifugal force	500g or 700g*		Up to 800g	Up to 2350g
Centrifugation timer	-		1 s – 99 min	1 s – 30 min
Vortexing mode	Single tube		Full rotor	
Vortexing intensity	Manual pressure		Soft, medium, hard	
Vortexing timer	-		0 – 20 s	
SMS-cycle regulation	-		1 – 999 cycles	
Tube protection	-	Stop at open lid		Lid lock
Display	-		LCD, 2x16 symbols	
Dimensions, mm	120x170x120	190x235x125		
Weight, kg**	1.4	1.7	2.1	2.5
Power consumption, W	30		11	24
Input voltage	230 V, 50 Hz or 120/230 V, 60 Hz		100-240 V, 50/60 Hz	
External power supply	-		DC 12 V	DC 24 V

The interested bidders are kindly requested to adhere to the above changes in order to submit their Quotations as per the UNDP requirement.

All other Terms and Conditions in this RFQ will remain in full force and unchanged.

Thank you in advance for your interest in UNDP procurement opportunities.

Sincerely

UNDP LIF Solicitations