

QUESTIONS AND ANSWERS REPORT

To: Bidders

From: UNDP Syria

Subject: Supply of Disability Aids and Prosthetic Devices & Equipment

Reference: UNDP-SYR-ITB-026-21

Date: 8-6-2021

UNDP Syria Procurement Unit received question regarding the subject RFP. This question is documented below with respective answers.

#	Questions & answers
	Item no. 4 Gel Liner with 4 Rings for Above Knee amputees
1	Which quantity is requested is it 100 pc as mentioned at Qty column or 125 as mentioned at Notes column:
	The total quantity is correct as 100 as shown in the quantity column . But we have mistaken in the disaggregated quantity per size the table in the "notes" column. Below is the correct quantities per size

below is the correct quantities per size			
Size	Circumference	QTY	
Α	26-27 cm	15	
В	28-31 cm	35	
С	32-37 cm	35	
D	38-43 cm	15	
	Total	100	

question about Item 1 lot 1 "SACH Foot" the requirement of removable inner part from the cover (shell)

We have adopted the generic scientific term of SACH foot covering many types and we are targeting non-wooden SACH models that allow for removing the inner part from the cover therefore the specification stay as it is.

3	question about the warranty period. Where some items in particular silicon liner are offered only with 6 months warranty from the manufacturer
	we agree to reduce the warranty period to 6 months to cover all items.
4	In the qualifications of the service providers, it is mentioned that we should "have a minimum of Seven years' experience in provision of similar services including employment, labour market governance, or relevant fields in the region preferably in Syria." Regarding Syrian-specific experience, we as a firm are planning to include local experts that would fit this criteria. Would this be satisfactory to UNDP, or are we expected to partner with a local organization instead?
	Both options are acceptable and satisfactory to UNDP.

All other terms and condition remain unchanged.